

## Research Article

**Cite this article:** Cringoli G *et al.* (2021). The Kubic FLOTAC microscope (KFM): a new compact digital microscope for helminth egg counts. *Parasitology* **148**, 427–434. <https://doi.org/10.1017/S003118202000219X>

Received: 18 July 2020  
Revised: 26 October 2020  
Accepted: 5 November 2020  
First published online: 20 November 2020


**Key words:**

Digital microscope; fecal egg count; Kubic FLOTAC microscope; Mini-FLOTAC; Tele-Parasitology

**Author for correspondence:**

Maria Paola Maurelli,  
E-mail: [mariapaola.maurelli@unina.it](mailto:mariapaola.maurelli@unina.it)

# The Kubic FLOTAC microscope (KFM): a new compact digital microscope for helminth egg counts

Giuseppe Cringoli<sup>1</sup>, Alessandra Amadesi<sup>1</sup>, Maria Paola Maurelli<sup>1</sup> , Biase Celano<sup>2</sup>, Gabriele Piantadosi<sup>3</sup>, Antonio Bosco<sup>1</sup>, Lavinia Ciuca<sup>1</sup>, Mario Cesarelli<sup>3</sup>, Paolo Bifulco<sup>3</sup>, Antonio Montresor<sup>4</sup> and Laura Rinaldi<sup>1</sup>

<sup>1</sup>Department of Veterinary Medicine and Animal Production, University of Naples Federico II, CREMOPAR, WHO Collaborating Centre ITA-116, Naples, Italy; <sup>2</sup>Officina Elettronica SPA S.r.l., Naples, Italy; <sup>3</sup>Department of Electrical Engineering and Information Technology, University of Naples Federico II, Naples, Italy and <sup>4</sup>Department of Control of Neglected Tropical Diseases, World Health Organization, Geneva, Switzerland

**Abstract**

The Kubic FLOTAC microscope (KFM) is a compact, low-cost, versatile and portable digital microscope designed to analyse fecal specimens prepared with Mini-FLOTAC or FLOTAC, in both field and laboratory settings. In this paper, we present the characteristics of the KFM along with its first validation for fecal egg count (FEC) of gastrointestinal nematodes (GINs) in cattle. For this latter purpose, a study was performed on 30 fecal samples from cattle experimentally infected by GINs to compare the performance of Mini-FLOTAC either using a traditional optical microscope (OM) or the KFM. The results of the comparison showed a substantial agreement (concordance correlation coefficient = 0.999), with a very low discrepancy ( $-0.425 \pm 7.370$ ) between the two microscopes. Moreover, the KFM captured images comparable with the view provided by the traditional OM. Therefore, the combination of sensitive, accurate, precise and standardized FEC techniques, as the Mini-FLOTAC, with a reliable automated system, will permit the real-time observation and quantification of parasitic structures, thanks also to artificial intelligence software, that is under development. For these reasons, the KFM is a promising tool for an accurate and efficient FEC to improve parasite diagnosis and to assist new generations of operators in veterinary and public health.

**Introduction**

Currently, the most widely used methods for diagnosis of intestinal parasites in both field and laboratory settings are fecal egg count (FEC) techniques, based on the microscopic identification and count of parasitic structures (eggs, larvae, oocysts and cysts) in fecal specimens, e.g. McMaster, Kato-Katz, FLOTAC, FECPAK, Mini-FLOTAC (MAFF, 1986; Cringoli *et al.*, 2010, 2017; Levecke *et al.*, 2011, 2012; WHO, 2019). Also, the availability of a reliable, low-cost, easy-to-use and quantitative test to perform the fecal egg count reduction test (FECRT) is of pivotal importance to facilitate the monitoring of deworming programmes in endemic countries (Cools *et al.*, 2019) and to determine anthelmintic efficacy/resistance in consideration of the growing concern of the emergence of anthelmintic resistance in humans (Vlaminck *et al.*, 2019), livestock (Vercruyse *et al.*, 2018; Kaplan, 2020) and pet animals (Jimenez Castro *et al.*, 2019; Kitchen *et al.*, 2019).

Among the above mentioned techniques, Mini-FLOTAC, in particular, is considered a good candidate for a standardized FEC/FECRT either in lab or in field setting (Nikolay *et al.*, 2014; Dias de Castro *et al.*, 2017; George *et al.*, 2017; Paras *et al.*, 2018; Cools *et al.*, 2019; Rinaldi *et al.*, 2019; Amadesi *et al.*, 2020). However, the Mini-FLOTAC, as all the other FEC techniques, requires specialized personnel and is time-consuming, especially when a large number of specimens are examined like in case of large epidemiological surveys (Cringoli *et al.*, 2017; Vercruyse *et al.*, 2018; Sukas *et al.*, 2019).

Recently, the use of new technologies is beginning to offer potential solutions to overcome gaps and limitations of FEC techniques (i.e. human errors and time for analysis). Recent studies have evaluated different smartphone-based technologies to magnify objects, to capture images or to perform an automated identification of endoparasites (i.e. protozoa and helminths), showing a good potential for wider application. Saeed and Jabbar (2018) reviewed the applications of various smartphone-based methods and devices developed from 1990 to 2017 for the diagnosis of different parasites of public health relevance (e.g. soil-transmitted helminths, *Schistosoma* spp., protozoa, etc.). Moreover, several studies, reported in Table 1, focused on the development of semi-automated and automated systems for assessing FEC in the veterinary field. However, these systems have shown difficulties in commercialization, mainly due to their low sensitivity and accuracy, high costs and/or the limited data on validations in the lab and/or in the field. For these reasons, we consider that there is still the need for a reliable automated system that improves the efficiency of parasitological diagnosis in veterinary medicine and public health at an affordable cost.

© The Author(s), 2020. Published by Cambridge University Press. This is an Open Access article, distributed under the terms of the Creative Commons Attribution-NonCommercial-NoDerivatives licence (<http://creativecommons.org/licenses/by-nc-nd/4.0/>), which permits non-commercial re-use, distribution, and reproduction in any medium, provided the original work is unaltered and is properly cited. The written permission of Cambridge University Press must be obtained for commercial re-use or in order to create a derivative work.

**Table 1.** Semi-automated and automated systems for parasite detection in humans and animals (principle, the hosts from which fecal samples were collected, the parasites detected, the advantages, limits and references)

Method	Principle	Hosts	Parasites	Advantages	Limitations	References
FECPAK <sup>G2</sup>	High-throughput technological system for <i>on-field</i> sample processing	Ruminants; humans	GINs; soil-transmitted helminths (STH)	Automated detection and count; remote parasite detection and data online management	Low sensitivity and accuracy	Mirams, (2016); Rashid <i>et al.</i> (2018); Moser <i>et al.</i> (2018); Vlamincck <i>et al.</i> (2018); Cools <i>et al.</i> (2019); Vlamincck <i>et al.</i> (2019)
Parasight system	Based on a fluorescent egg staining and a smartphone to capture images	Horses	Strongyles and <i>Parascaris equorum</i>	2.5 min test time, less variables and more accurate than McMaster technique	Validated only on horses	Slusarewicz <i>et al.</i> (2016); Scare <i>et al.</i> (2017); Cain <i>et al.</i> (2020)
Lab-on-disk platform	Based on a combined gravitational and centrifugal flotation and the use of a converging collection chamber to create a monolayer of eggs	Humans; pigs	STH, <i>Schistosoma mansoni</i> ; <i>Ascaris suum</i>	High-quality of images, permitting a good identification and count	High cost; potentially limited use in field, due to: (i) the need for a minicentrifuge; (ii) the need for electricity and (iii) relatively large size/difficult to carry over distance	Sukas <i>et al.</i> (2019)
Automated robotic system	Based on an automated X-Y stage; the autofocusing and scanning function are based on LabVIEW GUI	Monkey; dogs; sheep; cattle	<i>Trichuris</i> spp., <i>Toxocara</i> spp., strongyle, <i>Isospora</i> spp., <i>Eimeria</i> spp.	Inexpensive (~US\$350), compact, possibility to use fluorescence	The system can be used only with McMaster chamber; not validated	Lu <i>et al.</i> (2018); Li <i>et al.</i> (2019)
Automated diagnosis of intestinal parasites (DAPI)	Based on a motorized system to read slides, using a digital camera and machine learning software	Dogs	<i>Ancylostoma</i> spp., <i>Toxocara</i> spp., <i>Trichuris</i> spp., and <i>Giardia</i> spp.	Automated detection of eggs through machine learning software	High cost; not portable; not validated	Inácio <i>et al.</i> (2020)
Telenostic system	Automated digital microscope with a 10× lens using machine learning software	Cattle	GINs	High level of agreement between the prototype and manual systems of FEC (i.e. McMaster and Mini-FLOTAC)	Validated only on cattle; it requires long time to acquire and analyse the images (approximately 42 min)	Elghryani <i>et al.</i> (2020)
VETSCAN IMAGYST	Composed of a digital slide scanner and machine learning software	Dogs and cats	Ancylostomidae, <i>Toxocara</i> spp., <i>Trichuris</i> spp., Taeniidae	The system allows identification and count of eggs within 15 min	High cost; not portable; validated only on dogs and cats	Nagamori <i>et al.</i> (2020)



**Fig. 1.** KFM used with the Mini-FLOTAC (A) and FLOTAC devices (B).

In this paper, we present the characteristics of the Kubic FLOTAC microscope (KFM), a new compact, cubic in shape ( $20 \times 20 \times 20$  cm), low-cost (~600 euros), versatile and portable digital microscope capable of scanning fecal specimens prepared with Mini-FLOTAC or FLOTAC (Mini-FLOTAC/FLOTAC) (Fig. 1A and B). The KFM can be used also directly in field, because it does not need any additional special requirement (e.g. centrifuge or any electricity source), being provided of a lithium battery with an autonomy up to 20 h. Moreover, we report the findings of the first validation of the KFM to perform FEC of gastrointestinal nematodes (GINs) in cattle using the Mini-FLOTAC technique and compare the results obtained by KFM with those obtained by a traditional optical microscope (OM).

## Materials and methods

### *The Kubic FLOTAC microscope (KFM) description*

The KFM is composed of an electromechanical part that allows a three-dimensional (3D) scan of the Mini-FLOTAC/FLOTAC (reading discs) and software that allows remote interactions and digital image processing supported by artificial intelligence (AI) for the recognition of helminth eggs and other parasitic structures. FreeCAD (open source, LGPL license) and Design Spark Mechanical (RS-components, Corby, UK) software were used to perform the 3D mechanical of the KFM (Fig. 2A and B). Once opportunely prepared, you can insert the Mini-FLOTAC or FLOTAC devices into a specific slide-out tray of the KFM (similarly to inserting a DVD into a player). Then, the tray is withdrawn inside and specific 3D landmarks corresponding to the corners of the two flotation chambers of the Mini-FLOTAC/FLOTAC are automatically located. The KFM scanning device is equivalent to a XYZ motorized stage for microscopy.


The 3D positioning system of the motorized stage is based on a simple, non-standard Cartesian motor system solution and is provided with open-loop stepper motors coupled with precision

translation stages to achieve accurate 3D motion control (Fig. 2C and D). A standard, low-cost stepper motor driver based on Arduino nano-board 105 (Arduino AG, Ivrea, Italy) and a free Grbl firmware were adopted to remote control of the KFM.


The KFM optical part is mainly composed of an LED light source with a condenser, which provides brightness adjustment, and a digital camera, which permits adjustable magnification at 100 $\times$ , 200 $\times$ , 300 $\times$ , with a maximum resolution of 8 MPixel (3264  $\times$  2448 pixel) and the size of image of 1024  $\times$  768 (0.8 Mp) (Fig. 3). A Raspberry Pi (Raspberry Pi Foundation, Caldecote, UK) permits the remote control of the optical part. The motorized stage can automatically move stepwise to entirely scan the two flotation chambers of the Mini-FLOTAC/FLOTAC. For each step, the KFM camera captures a picture (or a Z-stack). Eventually, you can photograph the whole surface of the flotation chambers (each 18 mm by 18 mm wide) and the software stitches the multiple photos to obtain only one picture to perform the visual recognizing and the count of parasitic structures.

The KFM system has various options for external connectivity: (i) two USB ports to easily transfer the captured images; (ii) an Ethernet cable connector, a Bluetooth and a Wi-Fi wireless connection, which permit all users to easily transfer, share and backup the captured images to other devices. In addition, you can fully control the KFM using software by an external device, i.e. a smartphone, a tablet or a PC. A web interface permits the remote control of the KFM (Fig. 4); the internal software works on Linux operating system. Such an interface can be activated using any web browser. The remote interface supports Android 4.1, iOS 8.0, Windows 7/8/10 and MacOS X 10.8. We developed a specific app that allows the remote control of the KFM directly from mobile phones.

The software has two modalities: 'Administrator', password-protected, for the management of settings and mechanical movements and 'Viewer' to allow remote viewing. The images captured by the KFM can be stored in an Internet cloud and/or transmitted to a diagnostic hub for telediagnosis or parasitological consultation.



**Fig. 2.** (A) FreeCAD and Design Spark Mechanical of the KFM (external view); (B) FreeCAD and Design Spark Mechanical of the KFM (internal view); (C) schematic diagram showing mechanical, electronic and optical systems of the KFM and (D) a particular of the handling motor based on the no-standard Cartesian system.




**Fig. 3.** Digital imaging of GIN eggs (green arrow) and air bubbles (red square) using the KFM with a digital zoom 100 $\times$  (A), 200 $\times$  (B) and 300 $\times$  (C).

At the moment, for these characteristics, you can use the KFM as a digital microscope, instead of an OM, remotely controlled to perform the visual identification and counting of parasitic structures, as validated in this paper. Moreover, further developments are ongoing to update a fully automated version of KFM, as reported in the 'Discussion' section.

#### Study sampling and laboratory activities

From September to December 2019, 30 fecal samples were collected from Belgian Blue and Holstein cattle (6 months old) experimentally infected with 50 000 third stage larvae (L3) of *Cooperia oncophora* ( $n = 15$  calves) or *Ostertagia ostertagi* ( $n = 15$  calves) stabled at the experimental farm of the Faculty of



**Fig. 4.** Image of a part of a Mini-FLOTAC chamber, captured by technological devices (smartphone, tablet or a PC) connected with the microscope that shows GIN eggs (green arrows) and air bubbles (red squares).

Veterinary Medicine, Ghent University (Belgium). Then, we transferred the collected samples to the Department of Virology, Parasitology and Immunology of Ghent University. In the lab, an operator analysed six sub-samples for each fecal sample, filling six Mini-FLOTACs with 2 g of feces and homogenizing them with 38 mL (dilution ratio 1:20) of a saturated sodium chloride flotation solution (specific gravity = 1200). Then, six Mini-FLOTAC devices were filled (one per each sub-sample), following the SOPs described in Cringoli *et al.* (2017). To ensure the quality of parasitological examination, one operator read the Mini-FLOTACs with an OM (Leica Microsystems, Wetzlar, Germany), then a second operator randomized the Mini-FLOTACs, before the reading with the KFM to obtain blinded results. Finally, the total time was evaluated to perform the Mini-FLOTAC technique either with the reading under the KFM or under a traditional OM.

### Statistical analysis

A Mann–Whitney test was used to compare the readings by both microscopes. Moreover, we calculated the Lin's concordance correlation coefficient (CCC) and the corresponding 95% confidence interval (CI) to quantify the agreement between the analysis using the OM and the KFM. The agreement was classified as poor,

**Table 2.** Number of counted eggs (sum and mean) for the analysis performed by the Mini-FLOTAC using a traditional OM and the KFM at low (<10), medium (10–25), high (>25) egg count levels and total counts


Egg level	OM		KFM	
	Sum	Mean	Sum	Mean
Low	296	4.93	291	4.85
Medium	834	13.90	827	13.78
High	1761	29.35	1771	29.52
Total counts	2891	16.06	2889	16.05

moderate, substantial or almost perfect for CCC values <0.9, 0.90–0.95, 0.95–0.99 or >0.99, respectively (McBride, 2005). Finally, the level of agreement was analysed using the Bland–Altman plot. All the statistical analyses were performed using GraphPad Prism v.8 (Graph Pad Software, San Diego, CA, USA) and SPSS Statistics v.23 (IBM, Armonk, NY, USA). All tests were considered statistically significant at  $P < 0.05$ .

### Results

In total, 180 counts were performed using the OM and 180 counts using the KFM. A range of GIN eggs from 1 to 62 was found for each sample analysed, as depicted in Fig. 5. Based on the counted eggs, each sample was assigned to one of the three egg count levels: low (<10), medium (10–25) and high (>25). Table 2 shows the sum and mean egg counts detected by the OM and the KFM at the different count levels. The correlation and the agreement between counted eggs by both microscopes for each level, and total counts are reported in Figs 6 and 7.

The results analysed by the Mann–Whitney test showed that there was not a statistically significant difference ( $P > 0.05$ ) between medians of counted eggs obtained with the two microscopes, for each egg count level. The CCC between the reading with OM and KFM was substantial for low (CCC = 0.984; 95% CI = 0.973–0.990), medium (CCC = 0.981; 95% CI = 0.968–0.988) and high (CCC = 0.998; 95% CI = 0.996–0.999) egg count levels; similar findings were found when considering total egg counts (CCC = 0.999;



**Fig. 5.** Comparison of the mean of counted eggs between the readings by the OM and the KFM for each sample.




Fig. 6. Correlation between the number of counted eggs based on the examination of Mini-FLOTAC using the OM and KFM for each level of egg counts and total counts.




Fig. 7. Bland-Altman plot of number of counted GIN eggs based on the examination of Mini-FLOTAC using OM and KFM.

95% CI = 0.998–0.999) (Fig. 6). Furthermore, the Bland–Altman plot (Fig. 7) demonstrated an excellent agreement between the two reading approaches (bias =  $-0.425 \pm 7.370$ , with 95% limits

Table 3. Time of analysis to perform the Mini-FLOTAC technique with the traditional OM and the KFM for each step of the protocol used

Step	OM	KFM
Preparation of the sample using Fill-FLOTAC and Mini-FLOTAC	11 min	11 min
Reading of the Mini-FLOTAC	1–5 min	3–8 min
Analysis of results	5 s	5 s
Total	12–16 min and 5 s	14–19 min and 5 s

of agreement from  $-14.872$  to  $14.021$ ). Finally, the time required to perform the Mini-FLOTAC technique with the KFM and the traditional OM is reported in Table 3.

**Discussion**

The KFM presented for the first time in this paper has proven a promising system for an accurate assessment of GIN egg counts in

cattle. The results showed that there were not statistically significant differences between medians of counted eggs obtained with OM and KFM. Moreover, a substantial agreement (CCC = 0.999) and a low ( $-0.425 \pm 7.370$ ) discrepancy were found between the two microscopes. These findings were confirmed also by the Bland–Altman plot that was very useful to compare the egg counts obtained with the two microscopes. The Bland–Altman plot showed a very low bias ( $-0.425$ ) and the absence of a systematic error, in fact points of the graph corresponding to the differences between the readings with the two systems are all around the line zero, showing a dispersion of the points only for the low egg count level (<10 eggs), without outliers.

The time taken to scan and count the GIN eggs in the Mini-FLOTAC chambers under the KFM was  $\sim 3$ – $8$  min compared to  $\sim 1$ – $5$  min taken to read the Mini-FLOTAC under the traditional OM by the same skilled operator. However, the KFM system used in this comparison study was only a prototype, but we are developing an App that will allow us to improve the speed of scanning of the Mini-FLOTAC apparatus, thus reducing the time for analysis.

There are many advantages of the KFM compared to the OM, e.g.: (i) it can be used without requiring electricity in both laboratory and field settings; (ii) it is cheap ( $\sim \text{€}600$ ) and (iii) it is able to transfer *via* internet the captured pictures to other laboratories or can be used by remote from another country. Therefore, it could be very useful to create a network of laboratories or to support operators directly in the field as expected by Tele-Medicine and Tele-Parasitology (Di Cerbo *et al.*, 2015; Scheild *et al.*, 2007; Zaffarano *et al.*, 2018).

Moreover, this system is very easy to be performed by any operator without a specific training (it will be possible to use the KFM just following the manufacturer's instructions).

One of the future improvements of the KFM, actually under development, is the tuning of a predictive model, for the automated identification and counting of helminth eggs, based on AI (machine learning) using a single deep neural network, named Single Shot Multibox Detector (SSD) (Liu *et al.*, 2016). For this aim, the use of high-quality images, as well as those captured by the KFM in our study, is fundamental. To date, to perform the first ground-truth phase of the training software a dataset with 11 136 images of GIN eggs (confirmed by experts) from large and small ruminant feces was used. The acquired images are first elaborated to separate the parasitic structures from the background, which can be either impurities or pseudo-parasites, through the use of ImageJ (National Institute of Health) with the main goal of being able to characterize the parasite species and to automatically count eggs (Grishagin, 2015).

In future, this software will be available also for other parasitic structures of veterinary and public health importance.

Therefore, KFM permits the combination of a sensitive, accurate, precise and standardized FEC techniques, as the Mini-FLOTAC/FLOTAC (Cringoli *et al.*, 2010, 2017), with a reliable system to capture and analyse pictures. Moreover, thanks to the development of AI software for automated identification and counts, this system will allow a reduction in human errors and time of reading, increasing its diagnostic efficiency.

## Conclusions

In conclusion, the KFM is a promising system that may, upon successful validation studies performed in different settings and laboratories of medical and veterinary parasitology, allow progression in overcoming the numerous limitations of traditional egg counting techniques.

The development of the AI predictive model will be very useful for an easy-to-use, low cost and precise automated system for

identification and counting of parasitic structures allowing a rapid assessment of FEC/FECR to assist a new generation of operators (i.e. technicians, physicians, veterinarians and farmers) in veterinary and human parasitology and diagnostics. In principle, the KFM could also be adapted to read other fecal preparation slides such as the Kato-Katz or McMaster.

## Data

All data generated or analysed during this study are included in this published paper. The datasets used and/or analysed during the current study are available from the corresponding author upon reasonable request.

**Acknowledgements.** The authors acknowledge M.E. Morgoglione, M. Santaniello, P. Pepe, D. Ianniello, S. Pennacchio and V. Musella for their participation in the development of the Kubic FLOTAC microscope. Moreover, they would like to express sincere appreciation to Prof. E. Claerebout, Prof. B. Levecke, N. De Wilde, S. Casaert and all the Department of Virology, Parasitology and Immunology of the Faculty of Veterinary Medicine, Ghent University, Belgium, for their collaboration.

**Author contribution.** Invented the KFM: GC; all the authors participated in the KFM development. Conceived, designed and coordinated the validation study of the KFM to perform FEC of GINs in cattle: GC, LR and MPM. Performed sampling and laboratory analyses: AA. Performed the first ground-truth phase of training software: GP and AA. All authors contributed to data analysis and interpretation, and preparation of the manuscript. All authors read and approved the final manuscript.

**Financial support.** This research did not receive any specific grant from funding agencies in the public, commercial or not-for-profit sectors.

**Conflict of interest.** The Mini-FLOTAC technique was developed and is patented by GC, but the patent has been handed over to the University of Naples Federico II. The fact that GC is the current patent holder of the Mini-FLOTAC and Fill-FLOTAC had no role in the preparation and submission of the protocols reported or the design and implementation of ongoing and future studies. To obtain Mini-FLOTAC or Fill-FLOTAC devices, a contribution is required that is used only to cover costs of production and packaging, and to contribute to the ongoing FLOTAC research. The remaining authors declare that they have no competing interests.

## References

- Amadesi A, Bosco A, Rinaldi L, Cringoli G, Claerebout E and Maurelli MP (2020) Cattle gastrointestinal nematode egg-spiked faecal samples: high recovery rates using the Mini-FLOTAC technique. *Parasites & Vectors* **6**, 230.
- Cain JL, Slusarewicz P, Rutledge MH, McVey MR, Wielgus KM, Zynda KM, Wehling LM, Scare JA, Steuer AE and Nielsen MK (2020) Diagnostic performance of McMaster, Wisconsin, and automated egg counting techniques for enumeration of equine strongyle eggs in fecal samples. *Veterinary Parasitology* **284**, 109–117.
- Cools P, Vlaminck J, Albonico M, Ame S, Ayana M, José Antonio BP, Cringoli G, Dana D, Keiser J, Maurelli MP, Maya C, Matoso LF, Montesor A, Mekonnen Z, Mirams G, Corrêa-Oliveira R, Pinto SA, Rinaldi L, Sayasone S, Thomas E, Verweij JJ, Vercruyse J and Levecke B (2019) Diagnostic performance of a single and duplicate Kato-Katz, Mini-FLOTAC, FECPAKG2 and qPCR for the detection and quantification of soil-transmitted helminths in three endemic countries. *PLoS Neglected Tropical Diseases* **13**, e0007446.
- Cringoli G, Rinaldi L, Maurelli MP and Utzinger J (2010) FLOTAC: new multivalent techniques for qualitative and quantitative copromicroscopic diagnosis of parasites in animals and humans. *Nature Protocols* **5**, 503–515.
- Cringoli G, Maurelli MP, Levecke B, Bosco A, Vercruyse J, Utzinger J and Rinaldi L (2017) The Mini-FLOTAC technique for the diagnosis of helminth and protozoan infections in humans and animals. *Nature Protocols* **12**, 1723–1732.
- Dias de Castro LL, Abrahao CLH, Buzatti A, Molento MB, Bastianetto E, Rodrigues DS, Lopes LB, Silva XM, Freitas MG, Conde MH and de

- Almeida Borgese F (2017) Comparison of McMaster and mini-FLOTAC fecal egg counting techniques in cattle and horses. *Veterinary Parasitology: Regional Studies and Reports* **10**, 132–135.
- Di Cerbo A, Morales-Medina JC, Palmieri B and Ianniti T (2015) Narrative review of Tele-Medicine consultation in medical practice. *Patient Preference and Adherence* **9**, 65–75.
- Elghryani N, Crispell J, Ebrahimi R, Krivoruchko M, Lobaskin V, McOwan T, O'Connor W, Power E, Voisin B, Scholz D and de Waal T (2020) Preliminary evaluation of a novel, fully automated, Telenostic device for rapid field-diagnosis of cattle parasites. *Parasitology* **24**, 1–5.
- George MM, Paras KL, Howell SB and Kaplan RM (2017) Utilization of composite fecal samples for detection of anthelmintic resistance in gastrointestinal nematodes of cattle. *Veterinary Parasitology* **240**, 24–29.
- Grishagin IV (2015) Automatic cell counting with ImageJ. *Analytical Biochemistry* **473**, 63–65.
- Inácio SV, Gomes JF, Falcão AX, Celso Suzuki TN, Nagata WB, Loiola SHN, dos Santos BM, Soares FA, Laryssa Rosa SL, Baptista CB, Alves GB and Bresciani KDS (2020) Automated diagnosis of canine gastrointestinal parasites using image analysis. *Pathogens (Basel, Switzerland)* **9**, 139.
- Jimenez Castro PD, Howell SB, Schaefer JJ, Avramenko RW, Gilleard JS and Kaplan RM (2019) Multiple drug resistance in the canine hookworm *Ancylostoma caninum*: an emerging threat? *Parasites & Vectors* **12**, 576.
- Kaplan RM (2020) Biology, epidemiology, diagnosis and management of anthelmintic resistance in gastrointestinal nematodes of livestock. *Veterinary Clinics of North America: Food Animal Practice* **36**, 17–30.
- Kitchen S, Ratnappan R, Han S, Leasure C, Grill E, Iqbal Z, Granger O, O'Halloran DM and Hawdon JM (2019) Isolation and characterization of a naturally occurring multidrug-resistant strain of the canine hookworm, *Ancylostoma caninum*. *International Journal for Parasitology* **49**, 397–406.
- Levecke B, Behnke JM, Ajjampur SS, Albonico M, Ame SM, Charlier J, Geiger SM, Hoa NT, Kamwa Ngassam RI, Kotze AC, McCarthy JS, Montresor A, Periago MV, Roy S, Tchuem Tchuente LA, Thach DT and Vercruyse J (2011) A comparison of the sensitivity and fecal egg counts of the McMaster egg counting and Kato-Katz thick smear methods. *PLoS Neglected Tropical Diseases* **5**, e1201.
- Levecke B, Dobson RJ, Speybroeck N, Vercruyse J and Charlier J (2012) Novel insights in the faecal egg count reduction test for monitoring drug efficacy against gastrointestinal nematodes of veterinary importance. *Veterinary Parasitology* **188**, 391–396.
- Li Y, Zheng R, Wu Y, Chu K, Xu Q, Sun M and Smith ZJ (2019) A low-cost, automated parasite diagnostic system Via A portable, robotic microscope and deep learning. *Journal of Biophotonics* **12**, e201800410.
- Liu W, Anguelov D, Erhan D, Szegedy C, Reed S, Fu CY and Berg AC (2016) SSD: Single Shot Multibox Detector. In Liebe B, Matas J, Sebe N and Welling M (eds), *Computer Vision - ECCV 2016*. ECCV 2016. *Lecture Notes in Computer Science*, Vol. 9905. Cham: Springer.
- Lu Q, Liu G, Xiao C, Hu C, Zhang S, Xu RX, Chu K, Xu Q and Smith ZJ (2018) A modular, open-source, slide-scanning microscope for diagnostic applications in resource-constrained settings. *PLoS ONE* **13**, e0194063. doi: 10.1371/journal.pone.0194063.
- MAFF (1986) *Manual of Veterinary Parasitological Techniques*, 3rd Edn. London: Her Majesty's Stationary Office. p. 160.
- McBride GB (2005) A proposal for strength-of-agreement criteria for Lin's concordance correlation coefficient. Hamilton: National Institute of Water and Atmospheric Research Client Report: HAM2005-062.
- Miram G (2016) FECPAKG2 Is a novel online faecal egg counting system for either remote location or laboratory/clinic based parasite FEC diagnostics. *Proceedings of the Society of Sheep and Beef Cattle Veterinarians of the New Zealand Veterinary Association*, 65–69.
- Moser W, Bärenbold O, Mirams GJ, Cools P, Vlaminck J, Ali SM, Ame SM, Hattendorf J, Vounatsou P, Levecke B and Keiser J (2018) Diagnostic comparison between FECPAKG2 and the Kato-Katz method for analyzing soil-transmitted helminth eggs in stool. *PLoS Neglected Tropical Diseases* **12**, e0006562. doi: 10.1371/journal.pntd.0006562.
- Nagamori Y, Hall Sedlak R, DeRosa A, Pullins A, Cree T, Loenser M, Larson BS, Richard Boyd Smith RB and Richard Goldstein R (2020) Evaluation of the VETSCAN IMAGYST: an in-clinic canine and feline fecal parasite detection system integrated with a deep learning algorithm. *Parasites & Vectors* **13**, 346.
- Nikolay B, Brooker SJ and Pullan RL (2014) Sensitivity of diagnostic tests for human soil-transmitted helminth infections: a meta-analysis in the absence of a true gold standard. *International Journal for Parasitology* **44**, 765–774.
- Paras KL, George MM, Vidyashankar AN and Kaplan RM (2018) Comparison of fecal egg counting methods in four livestock species. *Veterinary Parasitology* **257**, 21–27.
- Rashid MH, Stevenson MA, Waenga S, Mirams G, Campbell AJD, Vaughan JL and Jabbar A (2018) Comparison of McMaster and FECPAKG2 methods for counting nematode eggs in the faeces of alpacas. *Parasites & Vectors* **2**, 278.
- Rinaldi L, Amadesi A, Dufourd E, Bosco A, Gadanho M, Lehebel A, Paola Maurelli M, Chauvin A, Charlier J, Cringoli G, Ravinet N and Chartier C (2019) Rapid assessment of faecal egg count and faecal egg count reduction through composite sampling in cattle. *Parasites & Vectors* **12**, 353.
- Saeed MA and Jabbar A (2018) 'Smart diagnosis' of parasitic diseases by use of smartphones. *Journal of Clinical Microbiology* **56**, e01469–17.
- Scare JA, Slusarewicz P, Noel ML, Wielgus KM and Nielsen MK (2017) Evaluation of accuracy and precision of a smartphone based automated parasite egg counting system in comparison to the McMaster and Mini-FLOTAC methods. *Veterinary Parasitology* **247**, 85–92.
- Scheidt P, Lam DM, Thömmes A and Zöller L (2007) Telemicrobiology: a novel Tele-Medicine capability for mission support in the field of infectious medicine. *Telemedicine Journal and e-Health* **13**, 108–117.
- Shot MultiBox Detector. ECCV. *Lecture Notes in Computer Science*, vol. 9905. Cham: Springer.
- Slusarewicz P, Pagano S, Mills C, Popa G, Chow KM, Mendenhall M, Rodgers DW and Nielsen MK (2016) Automated parasite faecal egg counting using fluorescence labelling, smartphone image capture and computational image analysis. *International Journal for Parasitology* **46**, 485–493.
- Sukas S, Van Dorst B, Kryj A, Lagatie O, De Malsche W and Stuyver LJ (2019) Development of a lab-on-a-disk platform with digital imaging for identification and counting of parasite eggs in human and animal stool. *Micromachines* **10**, 852.
- Vercruyse J, Charlier J, Van Dijk J, Morgan ER, Geary T, von Samson-Himmelstjerna G and Claerebout E (2018) Control of helminth ruminant infections by 2030. *Parasitology* **145**, 1655–1664.
- Vlaminck J, Cools P, Albonico M, Ame S, Ayana M, Bethony J, Cringoli G, Dana D, Keiser J, Maurelli MP, Montresor A, Mekonnen Z, Mirams G, Corrêa-Oliveira R, Prichard R, Rashwan N, Rinaldi L, Sayasone S, Thomas E, Verweij JJ, Vercruyse J and Levecke B (2018) Comprehensive evaluation of stool-based diagnostic methods and benzimidazole resistance markers to assess drug efficacy and detect the emergence of anthelmintic resistance: a Starworms study protocol. *PLoS Neglected Tropical Diseases* **12**, e0006912.
- Vlaminck J, Cools P, Albonico M, Ame S, Ayana M, Cringoli G, Dana D, Keiser J, Maurelli MP, Matoso LF, Montresor A, Mekonnen Z, Mirams G, Corrêa-Oliveira R, Pinto SA, Rinaldi L, Sayasone S, Thomas E, Vercruyse J, Verweij JJ and Levecke B (2019) Therapeutic efficacy of albendazole against soil-transmitted helminthiasis in children measured by five diagnostic methods. *PLoS Neglected Tropical Diseases* **13**, e0007471.
- WHO (World Health Organization) (2019) *Bench Aids for the Diagnosis of Intestinal Parasites*, 2nd Edn. Geneva, Switzerland: WHO.
- Zaffarano GP, Morandi B, Menegotto A, Ostanello F and Poglayen G (2018) Parasitological experiences in wild ruminants of South African preserves. *Journal of Veterinary Medicine and Animal Health* **10**, 67–71.